

REVIEW



WILEY

Allergic contact dermatitis caused by glucose sensors and insulin pumps: A full review

Part 2. Case reports and case series, clinical features, patch test procedures, differentiation from irritant dermatitis, management of allergic patients and (proposed) legislation

Anton de Groot¹ | Emma M. van Oers² | Norbertus A. Ipenburg¹ | Thomas Rustemeyer¹

¹Dermato-Allergology and Occupational Dermatology, Amsterdam University Medical Centers, Amsterdam, The Netherlands

²Amsterdam University Medical Centers, Amsterdam, The Netherlands

Correspondence

Anton de Groot, Dermato-Allergology and Occupational Dermatology, Amsterdam University Medical Centers, Meibergdreef 9, 1105 AZ Amsterdam, The Netherlands.
Email: antondegroot@planet.nl

Abstract

During the past 8 years, a large number of reports have appeared on allergic contact dermatitis to glucose sensors and insulin pumps in paediatric and adult patients with type 1 diabetes mellitus. Isobornyl acrylate in one particular sensor sensitised many hundreds of (published) individuals, and many other allergens were discovered in a large number of sensors and pumps. Diagnostic procedures with patch tests proved very complicated, as manufacturers showed a serious lack of cooperation with dermatologists in providing information on the ingredients of their products and samples for patch testing. This two part article provides a full and detailed review of all aspects of the subject of allergic contact dermatitis to glucose sensors and insulin pumps. Part 1 provided a general introduction to sensors and pumps, a survey of the cutaneous adverse reactions that they have caused, a full account of the allergens in the diabetes devices and an overview of the glucose sensors and insulin pumps that have caused allergic contact dermatitis. This part 2 presents all published case reports and case series, clinical features of allergic contact dermatitis, patch test procedures, differentiation from irritant dermatitis, management of allergic patients and (proposed) legislation.

KEYWORDS

allergic contact dermatitis, colophonium, contact allergy, continuous subcutaneous insulin infusion, diabetes mellitus, diabetes medical device, glucose monitoring, glucose sensor, insulin pump, isobornyl acrylate, *N,N*-dimethylacrylamide

1 | INTRODUCTION

This is the second part of a full review of allergic contact dermatitis caused by glucose sensors and insulin pumps. Part 1 provided a

general introduction to sensors and pumps, a survey of the cutaneous adverse reactions that they have caused, a full account of the allergens in the diabetes devices and an overview of the glucose sensors and insulin pumps that have caused allergic contact dermatitis.¹

This is an open access article under the terms of the [Creative Commons Attribution-NonCommercial-NoDerivs](https://creativecommons.org/licenses/by-nc-nd/4.0/) License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

© 2024 The Author(s). *Contact Dermatitis* published by John Wiley & Sons Ltd.

This part 2 presents all published case reports and case series, clinical features of allergic contact dermatitis, patch test procedures, differentiation from irritant dermatitis, management of allergic patients and (proposed) legislation.

2 | ALLERGIC CONTACT DERMATITIS FROM INSULIN PUMPS

Insulin pumps have been used since the mid-1970's to continuously deliver insulin, mainly to patients with type 1 diabetes mellitus. The older devices consisted of a needle that was inserted into the skin and fixed with a plastic butterfly and an adhesive patch, and was connected to a pump by a plastic cannula/catheter/tube. Several cases of allergic contact dermatitis (ACD) to these early pumps, or actually their infusion sets, were reported between 1985 and 2001, caused by the infusion needle (nickel), glues to fixate the needle to the cannula or plastic butterfly (epoxy resin, acrylates), or materials of the cannula itself (methyl methacrylate). These early reports are presented in paragraph 2.1.

Reports of ACD from modern insulin pumps and infusion sets began to appear from 2018 on with a case series of 4 patients from France who developed ACD from isobornyl acrylate (IBOA) in the Omnipod insulin pump.² Two patients were sensitised by the device itself after 4 months and 1 year. The other 2 already suffered ACD after the first use of the pump; they had previously become allergic to IBOA by the use of the glucose sensor FreeStyle Libre (FSL), which at that time had already sensitised a large number of patients using this device.³ Since then, cases of allergic reactions have been reported to 8 different brands of insulin infusion sets and patch pumps (part 1, tab. 9). They are presented in paragraph 2.2. An overview of all allergic reactions to the pumps with brand, culprit allergen(s), number of patients and literature references is provided in part 1, paragraph 4.4 (tab. 9).

2.1 | Early reports of allergic contact dermatitis to insulin pumps

Cases of allergic contact dermatitis to insulin pumps, or actually to their infusion sets, were already reported in the period 1985–2001.^{4–11} They are presented in some detail in Supporting Information S1.

2.2 | Allergic contact dermatitis to modern insulin pumps

Compared to glucose sensors, the number of reported cases of sensitization to and ACD from modern insulin pumps and infusion sets is modest. Allergic reactions to the Omnipod have been reported in nearly 20 patients, some of who had previously become sensitised to its ingredient IBOA from the use of the FSL sensor.^{12,13} The other

pumps/infusion sets (MiniMed Quick-set, MiniMed Sure-T, mylife Ypsopump Orbit, Omnipod DASH pump and the TouchCare A6) have each caused only 1 or 2 cases of allergic reactions in which the culprit allergen was established. This could possibly be explained by the fact that pumps are removed after 2–3 days (and then a new application follows to another part of the skin), whereas sensors may remain attached to the skin for 10–14 days, strongly increasing the risk of sensitization. Another possible explanation is that in the conventional pumps, the adhesive patches are separated from the housing of the pump by a cannula. This would preclude any contamination with allergens migrating from the housings, as has been the case with the Omnipod pump and the FSL sensor.

Several case series of ACD from insulin pumps and infusion sets have been reported.^{14–16} They are presented in some detail in Supporting Information File 1. Case reports with 1 or 2 patients with ACD have also been published^{17–28}; their key data are presented in Supporting Information Table S1.

3 | ALLERGIC CONTACT DERMATITIS FROM GLUCOSE SENSORS

The first report in 2017 of allergic contact dermatitis from IBOA in FreeStyle Libre (FSL) sensor by investigators from Belgium and Sweden³ was followed by a multitude of similar reports from other centres and countries (part 1, paragraph 4.3.1). Other sensors that have caused ACD include Dexcom G4, Dexcom G6, Dexcom G7, Enlite and Guardian 4, of which Enlite has caused most reactions with colophonium (derivatives), IBOA, or both as culprit allergens.

Over 20 reports have presented case series of ACD to glucose sensors^{29–49}; these are presented in some detail in Supporting Information File 2. There are also multiple case reports with 1–3 allergic patients available^{13,14,17,18,20,21,23–28,33–35,42,46,50–74}; their key data are presented in Supporting Information Table S2.

4 | CLINICAL FEATURES OF ALLERGIC CONTACT DERMATITIS

Allergic reactions to diabetes devices usually manifest as dermatitis under the adhesive patch of the glucose sensor or the insulin infusion set/patch pump. As diabetes type 1, for which the devices are generally used, mostly starts in (early) childhood, many of the patients are children.^{3,16,20,29,33,39,42,46} In case reports of ACD to sensors and pumps, >60% of the patients were younger than 18 years (Tables S1 and S2). The fact that the devices were primarily reimbursed by health insurers for children may also have played a role in this.

The time from the start of using a diabetes device to first appearance of dermatitis has varied from a few weeks to several years; typically, dermatitis started after 5–7 months.^{3,13,14,16,33,35,36,37,41} The majority of these data relate to patients who became sensitised to IBOA in the FreeStyle Libre sensor. However, in the case reports, in

which over half of the patients had ACD to other sensors and to pumps, the median delay was similar with 5 months (Tables S1 and S2).

Thereafter, upon repeated contact, ACD usually emerges within one to a few days, but sometimes later.¹⁴ Repeated applications in sensitised individuals often result in stronger allergic reactions with subsequent exposures and shorter time lags until dermatitis appears.⁶⁴ Some patients who have been sensitised by the use of a sensor or pump and who start using an alternative device may develop ACD to the new device within 1–3 days, when the allergen to which they have been sensitised previously is also present in the new device. This has occurred frequently in patients primarily sensitised to IBOA in the FSL sensor (e.g., References 12,13,20,21,24,26). However, when a device contains very low levels of the chemical to which the patient is allergic, it may take up to a few weeks before clinical ACD develops.²¹

The clinical picture of allergic reactions to diabetes devices may depend on the nature of the allergen, its concentration in the contact surface of the device, the strength of the sensitization, the contact time with the skin, the number of contacts and the degree of pre-existing skin damage, for example, from irritant dermatitis. Allergic reactions often present as subacute dermatitis with (sometimes severe) itching, erythema, mild swelling and scaling with well-defined borders in the form of the device's adhesive patch.⁴³ However, acute ACD with erythema, oedema, vesicles, oozing and erosions are not infrequent.^{30,31,61,64,75} Bleeding, secondary infection,^{53,72} abscesses,⁵³ scarring, and burning, stinging and pain have also been reported.^{46,75,76} In a few individuals ulcers have been observed, sometimes referred to as 'burn wounds' by patients, which, given their diabetes status, may bear an additional risk of superinfection.⁷⁵ In one series of 70 patients, of who 63 had reactions to FSL and 7 to Enlite, 70% had severe dermatitis, defined as red, swollen, intensively itching, oozing, bleeding or blistering dermatitis at the sensor contact site, and 30% had mild dermatitis.³³

Generally, the dermatitis is limited to the application surface of the adhesive patches or expands slightly to surrounding skin, which also applies to the more severe reactions.

Dermatitis usually heals within a week or so after removal of the device, sometimes leaving (post-inflammatory) hyperpigmentation^{30,46} or (less frequently) hypopigmentation. Leukoderma at the site of previous ACD has also been observed.^{10,63} In one of these cases, a potential role for hydroquinone monomethyl ether, which is present in the IBOA materials to prevent spontaneous polymerisation and which acts as a very potent depigmenting agent, was suggested.⁶³ It has been claimed that contact dermatitis to glucose sensors 'usually' manifests as acquired leukoderma,⁷⁷ which seems unlikely: leukoderma from diabetes devices has been reported only rarely^{10,63} and in both cases appeared *after* dermatitis had developed.

Rare cases of 'contact dermatitis syndrome' with generalised itchy erythema, vesicles and pustules on both soles⁶⁵ and systemic contact dermatitis⁷² have been reported.

The clinical aspect of allergic reactions to early insulin pumps was usually somewhat different from the more recent ones, presenting

with dermatitis under the 'butterfly' of the infusion set^{5,7} or linear-shaped lesions from allergic reactions to the cannula.^{4,9} One patient allergic to nickel developed painful papulonodular dermatitis at the site of the inserted needle with an itchy vesicular eruption spreading over her abdomen.⁸ Another nickel-allergic individual developed itchy papules around the needle site increasing to form a plaque 10 cm in diameter, later with papules spreading over the abdomen and buttocks with intense itching.⁶ The latter may well have been a case of ACD and systemic allergic dermatitis from nickel entering the bloodstream from the needle.⁶ These early cases are presented in paragraph 2.1.

The consequences of becoming sensitised to acrylates in glucose sensors or insulin pumps other than problems in finding safe alternative devices have not been well investigated. A few patients sensitised from diabetes devices have also suffered allergic reactions to acrylate nail cosmetics.^{35,50} Some were already sensitised to these cosmetics before using diabetes devices.³⁵ A causal relationship between sensitization to a sensor or pump followed by allergic reactions to nail cosmetics has never been obvious, although it was suggested in one case.⁵⁰ This may well be explained by the fact that isobornyl acrylate, the acrylate responsible for by far most cases of allergic reactions to diabetes devices, has very limited tendency to cross-react to other acrylates and to methacrylates such as 2-hydroxyethyl methacrylate (HEMA), methacrylates being the usual allergens in acrylate nail cosmetics (part 1, paragraph 4.3.1).⁷⁸ One patient sensitised to IBOA later developed ACD from a disposable blood pressure cuff, which was found to contain IBOA and 2-phenoxyethyl acrylate.⁵⁹

5 | DIAGNOSING ALLERGIC CONTACT DERMATITIS FROM DIABETES DEVICES: PATCH TEST PROCEDURES

Positive patch tests to one or more ingredients known to be present in diabetes devices that have caused cutaneous reactions suspected to be ACD confirm the diagnosis. However, the diagnostic process is complicated by difficulties in getting adequate information on what the medical devices are made of, obtaining information on how to patch test new suspected allergens and the collection of adequate patch test materials.²⁹ Indeed, information on the chemicals present in the adhesive patches and the housings of the devices is usually unavailable and most investigators have noticed a serious lack of cooperation from the manufacturers when requesting specific data on the ingredients of their products.^{3,14,17,21,24,33,35,42,43,46,75,79} This makes targeted testing very difficult. As a consequence, most allergens could only be identified by performing gas chromatography–mass spectrometry analyses (GC–MS) of (ultrasonic bath) extracts of the devices (part 1, paragraph 4.2).³¹ Only a few centers have the laboratory facilities and skilled personnel to perform such laborious and costly analyses.

We suggest to patch test patients with suspected ACD from diabetes devices with the European baseline series (or any other baseline series as appropriate), a diabetes device screening series, a (meth)

acrylate series, an isocyanate series, a plastic and glue series and, whenever possible, a piece of the adhesive of the device(s) that caused the dermatitis.^{46,80} Patch test reactions first appearing after D4 are not infrequent, especially with acrylates, but also with isocyanates, and therefore a test reading after 1 week is absolutely necessary.^{13,29,30,80}

In Table 1 a proposal for a diabetes device screening series containing 24 chemicals is shown largely based on the list of allergens that have caused ACD by their presence in glucose sensors, insulin pumps, or both, as shown in part 1, paragraph 4.3.6. Twelve of these are commercially available from Chemotechnique, 9 from Allergeaze. The other 12 can be purchased from a chemical company

TABLE 1 A proposal for a diabetes device screening series.

Allergens commercially available for patch testing	Concentration ⁱ	Available from	
		Chemotechnique	Allergeaze
Abietic acid	10.0%	+	+
Butyl acrylate	0.1%	+	+
Butylated hydroxytoluene (BHT) ^c	2.0%	+	+
4,4'-Diaminodiphenylmethane ^{b,d}	0.5%	+	+
Ethyl acrylate ^e	0.1%	+	+
Ethyl 2-cyanoacrylate	10.0%	+ ^g	+ ^h
1,6-Hexanediol diacrylate (HDDA) ³	0.1%	+	
Hydroabietyl alcohol (Abitol) ^c	10.0%	+	+
Isobornyl acrylate (IBOA) ^a	0.1%	+	+ ^h
Isophorone diisocyanate (IPDI) ^b	1.0%	+	
Methyl methacrylate ^a	2.0%	+	+
Tripropylene glycol diacrylate ^a	0.1%	+	
Chemicals not commercially available for patch testing	Concentration ⁱ	Available from	
		Sigma-Aldrich	TCI chemicals
2,4-di- <i>tert</i> -Butylphenol	1%	+	+
Carboxyethyl acrylate	0.1%	+	
Dicyclohexylmethane-4,4'-diisocyanate (DMDI)	1%		+
<i>N,N</i> -Dimethylacrylamide (DMAA)	0.3%	+	+
Dipropylene glycol diacrylate	0.1%		+
Hydroquinone monomethyl ether (4-methoxyphenol)	1%	+	+
Hydroxycyclohexyl phenyl ketone (1-benzoyl cyclohexanol)	5%	+	+
Isobornyl acrylate	0.3%	+	+
Isobornyl methacrylate	2%	+	+
2,2'-Methylenebis(6- <i>tert</i> -butyl-4-methylphenol) monoacrylate (MBPA)	1.5%		+
4,4'-Methylene diphenyl diisocyanate (MDI) (diphenylmethane 4,4'-diisocyanate) ^c	0.5%	+	+
2-Phenoxyethyl acrylate ^f	0.1%	+	+
Phenoxy(ethyleneoxy)ethyl acrylate (PEEA) (Phenol, ethoxylated, esters with acrylic acid); PEEA contains monomers (2-phenoxyethyl acrylate), oligomers and polymers	0.1% ^j	+	+

Abbreviations: Allergeaze: test allergens available from www.smartpracticecanada.com and (not all) www.smartpracticeeurope.com; Chemotechnique, www.chemotechnique.se; Sigma-Aldrich: www.sigmaaldrich.com; TCI: TCI Chemicals, www.tcichemicals.com.

^aMay be present in a '(meth)acrylate series'.

^bMay be present in an 'isocyanate series'.

^cMay be present in a 'plastic and glue series'.

^dAlso included as a screening agent for MDI sensitization.^{45,81}

^eFrequently co-/cross-reacted with isobornyl acrylate in one large study.³⁸

^fMonomer of phenoxy(ethyleneoxy)ethyl acrylate (PEEA).

^g'Currently not available due to temporary lack of high-grade raw material' (www.chemotechnique.se, October 22, 2024).

^hNot available from www.smartpracticeeurope.com.

ⁱAll chemicals in petrolatum.

^jBecause of the presence of only 14% monomers (2-phenoxyethyl acrylate) this test concentration may be too low.

(e.g., Sigma-Aldrich, TCI chemicals, other companies) for preparing the test allergens. Table 1 shows these chemicals with their test concentration as used in literature and availability from two major chemical suppliers.

Of the chemicals in the list of allergens identified as causes of ACD from diabetes devices (part 1, paragraph 4.3.6) colophonium, epoxy resin and nickel are present in the baseline series. Some patients allergic to colophonium only react to the material tested at 60%.^{29,31} Patients sensitised to modified colophonium or -derivatives such as hydroabietyl alcohol may have negative or ? + reactions to colophonium 20% in the baseline series (part 1, paragraph 4.3.2),^{16,82,83} hence the suggested presence of abietic acid and hydroabietyl alcohol in de screening series.

Some of the allergens may already be present in a '(meth)acrylate series' or 'isocyanate series' used by the investigator. Included in the screening series are also some chemicals that have not caused ACD from their presence in diabetes devices; for these, there are other reasons for their presence in the series. Ethyl acrylate, for example, is the only acrylate that fairly often co-reacts/cross-reacts with IBOA (part 1, paragraph 4.3.1). 4,4'-Methylene diphenyl diisocyanate (MDI) has been identified by GC-MS in diabetes devices (part 1, tab S1 in the Supporting information) and 4,4'-diaminodiphenylmethane (MDA) is a good screening agent for MDI allergy.^{45,81} Hydroquinone monomethyl ether (4-methoxyphenol) is present as inhibitor of spontaneous polymerisation in many acrylate raw materials used for the in-house preparation of patch test materials and should therefore be patch tested separately. We suggest isobornyl methacrylate, which is a frequent constituent of nail cosmetics⁸⁴ to be added to investigate cross-reactivity from IBOA sensitization.⁴⁴ 2-Phenoxyethyl acrylate, finally, is the monomer in phenoxypoly(ethyleneoxy)ethyl acrylate (PEEA), which has caused allergic reactions in diabetes devices and in which 2-phenoxyethyl acrylate is present in a concentration of nearly 15% (part 1, paragraph 4.3.4).

The test concentrations for chemicals not available for patch testing are based on the most recent published data. IBOA and DMAA were first tested at 0.1%. Later, it was found that a higher concentration of 0.3% identified allergic patients who had negative patch tests to 0.1%.^{12,13,29,30,46} IBOA 0.3% was negative in 100 controls and no late-appearing reactions suggestive of patch test sensitization were recorded.¹³

2,2'-Methylenebis(6-*tert*-butyl-4-methylphenol) monoacrylate (MBPA) was first tested at 0.3% pet.,³⁰ but later 1.5% proved to identify more cases of sensitization without causing irritant reactions or late reactions in controls (part 1, paragraph 4.3.4).³¹ Thus, a screening series is not static but should regularly be updated when new information becomes available. The medical device series used since 2020 in Malmö provides additional allergens which may be suitable for inclusion in a diabetes screening series.¹³

It should be realised that the performance of non-commercially prepared patch test materials, at the concentrations shown below (based on literature data), depends on the properties of the material purchased from a chemical company. It cannot be guaranteed that the test materials are non-irritant, safe, and adequate or optimal for detecting contact allergy. In addition, the series contains a large

number of acrylates, which may in some patients result in fierce allergic reactions or false positives (irritant reactions), and the risk of active sensitization is unknown.

Therefore, the screening series, chemicals and the test concentrations shown in the table should not be considered as an *advice* from the authors. It is presented as an option for (inclusion in) a diabetes device screening series to be considered and critically evaluated by the readers. Also, prevailing and applicable legislation for use of such materials for investigating patients should always be taken into account.

Pieces of the device (notably the adhesive material(s)) can be tested 'as is'.⁴⁶ Unfortunately, the concentration of the culprit allergen is often very low, especially when the allergen has migrated into the patch from the housing, which frequently results in false-negative tests.³⁰ Prolonging the application time from 2 to 4 days may increase the sensitivity of the test.³¹

When the results of all patch tests are negative and the patient is strongly suspected of ACD, additional patch testing with ultrasonic bath extracts of the involved device is recommended.^{80,85,86} It is preferable, but usually impossible, to make extracts from several devices of the same type. To increase the likelihood of extracting both polar and non-polar sensitizers, three extracting solvents with different physicochemical properties can be used—water, ethanol, and acetone.⁸⁰ Most investigators have used acetone for extraction (which is suitable for both polar and non-polar chemicals), some methanol (part 1, paragraph 4.2). A test reading after 1 week is mandatory unless positive reactions are noted already at the reading of D3 or D4. To substantially diminish the possibility of a false-positive reaction, patch testing in 20 consecutive dermatitis patients is necessary and should be negative.⁸⁰ Following this, GC-MS analyses should preferably be performed to identify potential allergens in the device(s) and these must, whenever possible, subsequently be patch tested in the patient and, when positive, in 20 controls. Unfortunately, false-negative reactions to extracts also occur.³⁰ Extracts giving a positive reaction can also be used for separation and identification based on patch testing with thin layer chromatograms, although very likely few centers will have the expertise and possibilities for such analyses.^{29,80}

When patch testing reveals contact allergy to a chemical known to be present in the device or to the non-irritant extract, the diagnosis of ACD has been established. Depending on the patch test results and clinical data, authors from Malmö, Sweden (who have greatly contributed to the scientific literature on allergic reactions to diabetes devices) have recommended as alternative diagnoses ACD?, UCD (unspecified contact dermatitis) and ICD (irritant contact dermatitis). The suggested criteria for making these diagnoses can be found in Reference 80.

6 | NEGATIVE PATCH TESTS IN PATIENTS SUSPECTED OF ALLERGIC CONTACT DERMATITIS: IRRITANT CONTACT DERMATITIS?

Many patients suffering from diabetes mellitus who use a glucose sensor, insulin pump or both develop cutaneous reactions to one or more

of their devices, frequently dermatitis (part 1, paragraph 3). Initially, skin reactions to the newly developed glucose sensor FSL were considered to be irritant rather than allergic, caused by skin temperature, occlusion, humidity, and long exposure.^{87,88} However, in 2017 isobornyl acrylate was found to be an important allergen in FSL causing ACD in 12 of 15 patients (80%) with skin reactions to this sensor.³ Later, *N,N*-dimethylacrylamide was reported as another sensitizer in this sensor (part 1, paragraph 4.3.3).

Nevertheless, irritant contact dermatitis (ICD) is presumably the more common reaction pattern, which is less severe and less reproducible upon application at new sites than ACD. It may be assumed that only the patients with more severe and persistent symptoms suggestive of ACD are referred to a dermatologist for patch testing. Indeed, in a university hospital in Belgium, in 2019, only 5.5% of diabetes patients were referred to the patch test clinic for suspected ACD from the FSL sensor.³⁹ However, even in these selected patient groups, relevant contact allergies could not always be demonstrated and the question has been raised repeatedly whether such patients had ICD.^{17,35,39,43,44} In a university center in Brussels, Belgium, of 52 patients who were referred for skin reactions to diabetes medical devices, relevant contact allergy could not be established in 17 (33%).⁴³ In other studies with case series of patients suspected of ACD to sensors, pumps or both, percentages of individuals in who the suspected contact allergy could not be verified by patch tests have varied widely: 7%,³ 13%,¹³ 13%,⁴⁶ 20%,³³ 24%,¹⁶ 31%,⁴⁴ 32%,³⁹ 43%,⁴² 67%,³⁵ and even 80%.¹⁷

In the Belgian study previously mentioned, all 52 patients, both those with ($n = 35$) and without ($n = 17$) relevant patch test reactions, presented a very similar clinical picture with an annular, squamous, well-defined itchy erythema under the sensor. The authors concluded that 'the clinical aspect, therefore, does not allow a distinction between ICD and ACD'. Also, the time between first symptoms and first use of the sensor was virtually the same for patients with ACD and suspected ICD (6–7 months). Following the first development of dermatitis, in both groups, all patients described progressive worsening and a decrease in the delay between sensor applications and appearance of the reaction, which was, according to the authors, more likely in ACD than ICD. Nevertheless, ICD was suggested to be present in the 17 patients (33%) in who no relevant contact allergies were found (with undiagnosed contact allergy as another possibility).⁴³ Subsequently, Swedish investigators argued that ICD was an unlikely diagnosis in many of the reported patients in the Belgian study.⁸⁰ They pointed out that there had not been a D7 reading, that the test concentration of 0.1% for IBOA and DMAA fails to identify a number of sensitised patients and that the temporal relationship between exposure and dermatitis in all patients argues in favour of the same mechanism for contact dermatitis in both (ACD and ICD) groups.⁸⁰

Indeed, there are several possible explanations for not finding a relevant contact allergy in patients in who the cutaneous reaction to diabetes devices is strongly suspected to be ACD (sometimes termed 'false-negative reactions'):

1. Patch tests are read only at D2 and D3/4, but not at D7. Thus, late-appearing reactions at D7 (negative at D3/4), which are not infrequent with IBOA, DMAA and other acrylates,⁸⁰ are missed¹³;
2. The test concentration used is too low and misses some cases of sensitization. IBOA and DMAA have long been tested at 0.1% pet. which we now know fails to identify a number of sensitizations which are picked up with testing at 0.3%.^{12,13,29,30,46} Already in 2020 investigators from Sweden have advised to test these allergens at 0.3% pet.⁸⁰ However, it may be assumed that by far most centers still test with IBOA 0.1% pet., as this is the only IBOA test allergen currently available from the major commercial providers. On the same note: 2,2'-methylenebis(6-*tert*-butyl-4-methylphenol) monoacrylate (MBPA) was first tested at 0.3% pet.,³⁰ but later 1.5% proved to identify more cases of sensitization (part 1, paragraph 4.3.4)³¹;
3. One or more allergens known to be present in diabetes devices are not tested;
4. The allergic reaction is caused by an established allergen, the presence of which in diabetes devices is not known and which is not patch tested;
5. The allergic reaction is caused by a chemical in the device not previously described as an allergen (and therefore not tested);
6. No patch tests were performed with the device itself (usually the adhesive patch) tested 'as is', or extracts (mostly acetone), which might have been positive, establishing ACD after negative control testing^{3,20,38,46};
7. Patch tests were performed with pieces of the device, extracts or both, but were negative due to their low content of the allergen; such false-negative reactions are relatively frequent^{30,38};
8. The patient has a weak contact allergy, for which the usual application time of 2 days is too short to provoke the allergic response.³¹

It cannot be excluded, however, that some patients who are, based on patient history and clinical symptoms, strongly suspected of ACD, in fact have severe irritant contact dermatitis and therefore show no relevant contact allergy even when optimal patch testing has been performed. Patch test procedures in suspected cases of ACD from diabetes devices are discussed in paragraph 5.

7 | FREQUENCY OF ALLERGIC CONTACT DERMATITIS FROM DIABETES DEVICES

Few studies have investigated the frequency of ACD from diabetes devices in groups of patients using them, and some of these were flawed.^{20,33,39,42,89} In a university hospital in Antwerp, Belgium, in the period December 2016 to April 2019, of 1036 patients (614 adults, 422 children) using FSL, 34 of the adult patients (5.5%) and 23 of the children (5.5%), total group 57 (5.5%), were referred to the dermatology department because of cutaneous adverse reactions to the sensor. When patch tested, 39/57 patients had positive reactions to

IBOA 0.1% pet., establishing ACD to the FSL sensor. Thus, the frequency of ACD to the device was 3.8% (39/1036).³⁹ This percentage is almost certainly an underestimation, as there was probably no late (D7) reading, and neither IBOA 0.3% nor DMAA were tested (these data were probably not available at the time), which inevitable has resulted in some false-negative reactions (paragraph 6). Also, as ACD can take months to even years to develop, some FSL users from this cohort may have developed ACD after the termination of the study. And finally, very likely not all patients with cutaneous reactions were referred for patch testing.³⁹

In two university hospitals in Finland, between October 2017 and January 2019, 70 patients were patch tested because of suspected ACD from FSL ($n = 63$) or the Enlite sensor ($n = 7$).³³ 51 FSL users had positive reactions to IBOA 0.1%. In November 2018, there were 6567 FSL users in the 2 hospitals, and the frequency of ACD therefore can be calculated as $51:6567 = 0.8\%$ (incorrectly stated to be 0.7%). There were 4 cases of ACD to Enlite in 385 users, which amounts to 1.0% (incorrectly stated as 0.8%).²⁸ Obviously, for the same reasons as given above, the actual figures are almost certainly higher. In addition, only patients who had to stop using the sensor or needed to use skin protectors under the sensor were referred for patch testing, indicating that patients with less severe allergic reactions were not patch tested.³³

In a university hospital in Brussels, between January 2016 and July 2019, 215 children were followed-up for wearing a FSL sensor. Ten patients were patch tested because of cutaneous reactions to the sensor and 9 had a positive reaction to IBOA 0.1% pet. This means that 9/215 (4.2%) of the children who were wearing the glucose sensor FSL were sensitised to IBOA,²⁰ which is, again, highly likely an underestimation.

In a university hospital in Spain, between March 2019 and February 2020, all children ($n = 264$) with type I diabetes mellitus who were equipped with glucose sensors were investigated.⁴² The data presented showed that 6.9% of the 275 sensors caused ACD in 14 of the 264 children (5.3%). However, the article contained many mistakes and in fact, the percentage of allergic reactions to sensors was not 6.9% but 4% and the percentage of patients with ACD to a sensor was not 5.3% but 3% (for details see Supporting information [File 2](#), FreeStyle Libre 1, Spain, 2020).

An Italian study calculated/estimated the prevalence of ACD to diabetes devices at 8.4%, but the research (in which apparently no dermatologists were involved) contained many flaws and this percentage is almost certainly unreliable (Reference [89](#) partly also presented in Reference [90](#)).

8 | MANAGING ALLERGIC CONTACT DERMATITIS FROM DIABETES DEVICES

After removal of the device that causes the skin reaction, ACD will heal spontaneously. In patients with severe dermatitis, topical corticosteroids will speed up the healing process. Next, measures must be taken to avoid recurrences of dermatitis. For this, simply relocating

the culprit device is of no help in cases of ACD. Only complete avoidance of a culprit sensitizer or at least a substantial decrease in exposure to it is necessary to prevent progressively worsening and relapsing dermatitis.⁷⁵ Using an alternative device may solve the problem, but is risky, as it may well contain the same or cross-reacting allergens and using no diabetes devices anymore is the safest option. However, the use of glucose sensors and insulin pumps often considerably enhances the quality of life of the patients. Therefore, many of them are reluctant to stop using their device^{40,43} and will first look for measures to alleviate the symptoms or, ideally, prevent the development of dermatitis completely. To achieve this, protective measures (secondary prevention) can be taken in patients who want to continue to use the culprit device. Alternatively, the patient can switch to another device that does not contain the allergen. Both options have proven to be rather challenging.

8.1 | Secondary prevention

There are three main strategies (sometimes used in combination) for secondary prevention while continuing the use of devices that causes allergic reactions: topical application of corticosteroids, applying barrier sprays before each new application, and applying a plaster between the skin and the adhesive patch of the device to stop the allergens from reaching the skin. It is often cited that the UK government website in 2019 stated that applying barrier creams, patches and sprays might affect the performance of the device; however, zero evidence was provided.⁹¹ By far most experience has been gained with patients reacting to IBOA in FreeStyle Libre 1 (which is not used anymore). The results of most protective measures have been somewhat disappointing^{3,17,33,39} and sometimes protective dressing even aggravated the dermatitis.³ In one study, only 14 of 63 (22%) patients with ACD to their IBOA-containing glucose sensor were able to continue using the device, with all 14 requiring use of a barrier agent and still having residual dermatitis.³³

8.1.1 | Topical corticosteroids

Twelve paediatric patients using diabetes devices who had developed local skin reactions under the adhesive patch of their sensor were treated with fluticasone propionate aqueous nasal solution on the skin area prior to adhesion of a new device. Five of the patients with moderate-to-severe local reactions had undergone 'allergic assessment demonstrating a positive response to their adhesive material'. In 10/12, there was no recurrence of local irritation nor dermatitis during almost 6 months of treatment.⁹² It is uncertain how many of the children had ACD, possibly those who had undergone 'allergic assessment'.

Fluticasone propionate sprayed topically fortnightly at the site prior to application of the sensor led to a marked improvement in one patient.⁵² In another report, the use of topical corticosteroids had very limited protective effect in an unspecified number of patients.³³

8.1.2 | Barrier sprays

Many patients have tried film-forming barrier sprays to prevent the adverse cutaneous reactions under their diabetes devices. Most frequently used was Cavilon™ spray (3M, St. Paul, Minnesota, USA), which is commonly used for the prevention of incontinence-associated dermatitis and for ostomy care. Its ingredients are hexamethyldisiloxane, isooctane, acrylate terpolymer and polyphenylmethylsiloxane. Although there is a case report claiming success,⁶⁵ its protective power appears to be limited, as it shields the skin only for a few days, whereas sensors may remain in place for up to 2 weeks.⁶⁵

In a Belgian study 16 patients of who 14 showed ACD to IBOA, had tried Cavilon to prevent their cutaneous adverse reactions. Five of these 16 patients (31%), all children, reported some improvement with this product, but only 2 continued its use. The first child, sensitised to IBOA, reported residual but acceptable itch and erythema, whereas the other child, not sensitised to IBOA, became completely free of symptoms. The 11 of 16 remaining patients (71%), all adults, experienced no benefit from Cavilon at all.³⁹ In another report, the use of Cavilon also had limited protective effect in an unspecified number of patients.³³ It should also be realised that ACD and severe irritant contact reactions to Cavilon have been reported in patients using the product for stoma care.⁹³

8.1.3 | Plasters and dressings

Another possibility for trying to avoid relapses of ACD is to apply a physical barrier between the skin and the adhesive part of the sensor, for example, with sterile adhesive tapes (plasters) or hydrocolloid- or silicone-based dressings/plasters/plates. Examples of such dressing are Compeed™ sheets (HRA Pharma, Paris, France),^{33,69} Cutimed® Hydro B (BSN Medical, Hamburg, Germany),^{40,68} Eakin® Surround protectors (Eakin Healthcare, Comber, Ards, United Kingdom),³³ Hansaplast™ blister plasters (Beiersdorf, Hamburg, Germany),^{26,40,68} Opsite® (Smith & Nephew, London, UK),³ Stomahesive™ (Convatec Group plc, Paddington, London, UK),^{24,25,40,54} and Tegaderm™ (3M, Minneapolis, MN, USA).^{3,51,64,94}

In some case reports, protective success has been claimed for the use of Tegaderm,^{64,94} Stomahesive,^{24,25} Compeed,⁶⁹ and Hansaplast blister plaster.²⁶ In other case reports and case series, dressing barriers were unsuccessful in preventing development of dermatitis or had limited effect.^{2,13,51} Authors from Australia presented 3 patients with ACD from IBOA, who were given Stomahesive as a barrier. Two patients could continue using their sensor without further development of ACD and the third had 'improvement' from the combination Stomahesive and a topical corticosteroid.⁵⁴ In a study from Poland, 3 of 10 Eakin Surround users were able to continue using FSL with no dermatitis.³³ In Germany, good results have been achieved with placing two or three overlapping Hansaplast blisters between the skin and the sensor. One patient 'was almost always able to keep his FSL in place for the full application time of 14 days' (meaning that dermatitis

still developed) and in 9 or 10 other patients 'this worked well' in terms of 'mostly preventing ACD symptoms'. The authors also claimed success for Cutimed Hydro B in another 2 patients.⁶⁸ Seven of these patients (5 using Hansaplast and 2 Cutimed Hydro B) were presented in more detail by these authors in another publication, where they also claimed success with Stomahesive in one individual. The authors did mention, however, that the usage of these protective devices has some limitations.⁴⁰

Thus, it appears that the use of hydrocolloid- or silicone-based plates underneath the sensor may help a subgroup of patients to tolerate the diabetes devices to which they are allergic. However, many patients still experience dermatitis to some extent.⁷⁵ Unfortunately, such protective skin barriers make wearing the devices less comfortable.⁶⁸ In addition, their centres have to be perforated so that the sensor filament or infusion cannula can reach the skin. This work-around sometimes prevents correct placing of the device, which may hamper its proper functioning.⁶⁸ Occlusion and additional costs for the plates are other critical points.^{24,40,75} In addition, the patches have poorer resistance to water contact, requiring more frequent replacement of the sensor.⁵³ Finally the dressings, some of which may contain acrylates or (modified) colophonium, may well cause allergic reactions themselves.^{95,96}

Generally, using these plates is regarded only as a temporary solution.^{68,75} In the end, many patients, despite taking protective measures still have to give up on the use of their devices and look for an alternative device (paragraph 8.2).³⁶

8.2 | Choosing an alternative sensor or pump

Many patients who have become sensitised to their device will try to continue its use by taking measures to prevent recurrences of dermatitis or ameliorating the symptoms (paragraph 8). When this proves to be unsuccessful, the use of an alternative device may be considered. Such a new device should not contain the allergen that has caused ACD in the patient or chemicals that may cross-react with it. For individuals in who the allergen has not been identified (paragraph 6) there is no alternative other than the 'by trial and error' method. Unfortunately, finding a suitable replacement device is equally difficult for patients with an established contact allergy, as there is virtually no information available on the composition of currently commercialised glucose sensors and insulin pumps.^{13,16} Manufacturers do not need to disclose the chemical composition of their product⁹⁷ and most have shown a serious lack of cooperation when requested to provide specific data on the ingredients of their products.^{3,14,17,21,24,33,35,42,43,46,79,97}

There is a great deal of information available, collected between 2016 and now, on the presence (part 1, paragraph 4.2, tabs 3 and S1 in the supporting information) or absence (part 1, paragraph 4.2, tables S2 and S3 in the supporting information) of specific allergens in specific sensors and pumps from chemical analyses (GC-MS) or (in a small minority) from information provided by the manufacturer. Some of this information is now outdated. Devices such as FSL and Dexcom G4 platinum, for example, are not available anymore, only their

successors. Also, product ingredients may be altered without a change of brand name or notification to the clinicians or the users.³⁰ Therefore, the information formerly acquired, although providing some direction in suggesting an alternative device, may not be accurate anymore. In addition, a negative GC-MS analysis for a certain chemical may not always completely rule out its presence in the product investigated and a guarantee that every possible allergen is identified cannot be given either.⁹⁷

In clinics investigating patients with suspected ACD to diabetes devices, having updated and accurate knowledge on the composition of the products is necessary. Only then can recommendations on alternative products be given to patients with known allergy to protect them from re-exposure to their sensitizers. When no information is provided by manufacturers, the only option to acquire data on the composition of products is by chemical analyses of extracts, which should be repeated periodically because of possible changes in the product composition. However, chemical analyses of the products can only partly, and often with considerable delay, compensate for the lack of information provided by manufacturers on the composition of their products^{15,80} and few centres have the opportunity to perform these investigations. Only a detailed labelling of the composition of these medical devices enables patients with known ACD to one of the components to avoid further skin reactions (paragraph 9).

9 | LEGISLATION

For reasons discussed in paragraph 8.2, it is important if not imperative for medical professionals to have access to detailed product descriptions of medical devices including their qualitative chemical composition in order to be able to perform appropriate patch testing, give advice to patients, and provide alternatives in case of allergies. As manufacturers continue to refuse collaboration and communication, several dermatologists have called attention to the need for (mandatory) full ingredient labelling of diabetes devices and other medical devices, both for existing products and for new devices entering the market.^{16,43,46,74,80,97,98,99} An Editorial in 2019 clearly showing the need for EU legislation to require disclosure and labelling of the composition of medical devices⁹⁹ was followed in 2021 by a Position statement from 14 experts in contact allergy on behalf of the European Society of Contact Dermatitis (ESCD), European Environmental and Contact Dermatitis Research group (EECDRG), the European Academy of Dermatology and Venereology (EADV) Contact Dermatitis Task Force, and the European Academy of Allergy and Clinical Immunology (EAACI).⁹⁷ This position paper reviewed the scientific literature on allergic reactions to medical devices and the current regulatory framework adopted for medical devices. It showed the negative impact of the absence of product labelling on patient care, costs of illness to the patients and society, and on the quality of life of the patients who have become allergic to a device. The authors strongly recommend that the European Commission require full labelling for any device.⁹⁷ Furthermore it was encouraged to implement an obligation for manufacturers of medical devices to cooperate and

disclose all information necessary for the management of patients who have suffered an adverse event. This particularly concerns full disclosure of the components (to the treating physician) of the device that has induced the reaction in order⁹⁵ and providing samples for patch testing.¹⁶

A large proportion of patients with ACD from diabetes devices is children. Many have used several sensors, pumps or both and develop several contact allergies. This emphasises the need for collaboration between the medical profession, the patient organisations and the companies producing the items, since these individuals will both need to and want to continue using diabetes devices, very likely for the rest of their lives.²⁹

AUTHOR CONTRIBUTIONS

Anton de Groot: Conceptualization; investigation; visualization; project administration; writing – original draft; writing – review and editing. **Emma M. van Oers:** Conceptualization; investigation; visualization; writing – original draft; writing – review and editing. **Norbertus A. Ipenburg:** Formal analysis; visualization; writing – review and editing. **Thomas Rustemeyer:** Supervision; writing – review and editing.

CONFLICT OF INTEREST STATEMENT

The authors declare no conflicts of interest.

DATA AVAILABILITY STATEMENT

Data sharing is not applicable to this article as no new data were created or analyzed in this study.

ORCID

Anton de Groot  <https://orcid.org/0000-0002-6666-7292>

Norbertus A. Ipenburg  <https://orcid.org/0000-0003-0679-4131>

Thomas Rustemeyer  <https://orcid.org/0000-0001-7580-0684>

REFERENCES

- De Groot AC, van Oers EM, Ipenburg NA, Rustemeyer T. Allergic contact dermatitis caused by glucose sensors and insulin pumps: a full review. Part 1: sensors and pumps, adverse cutaneous reactions, allergens, and diabetes devices causing allergic contact dermatitis. *Contact Dermatitis*. 2024.
- Raison-Peyron N, Mowitz M, Bonardel N, Aerts O, Bruze M. Allergic contact dermatitis caused by isobornyl acrylate in OmniPod, an innovative tubeless insulin pump. *Contact Dermatitis*. 2018;79(2):76-80. doi:[10.1111/cod.12995](https://doi.org/10.1111/cod.12995)
- Herman A, Aerts O, Baeck M, et al. Allergic contact dermatitis caused by isobornyl acrylate in freestyle[®] libre, a newly introduced glucose sensor. *Contact Dermatitis*. 2017;77(6):367-373. doi:[10.1111/cod.12866](https://doi.org/10.1111/cod.12866)
- Boom BW, van Driel LM. Allergic contact dermatitis to epoxy resin in infusion sets of an insulin pump. *Contact Dermatitis*. 1985;12(5):280. doi:[10.1111/j.1600-0536.1985.tb01137.x](https://doi.org/10.1111/j.1600-0536.1985.tb01137.x)
- Busschots AM, Meuleman V, Poesen N, Dooms-Goossens A. Contact allergy to components of glue in insulin pump infusion sets. *Contact Dermatitis*. 1995;33(3):205-206. doi:[10.1111/j.1600-0536.1995.tb00554.x](https://doi.org/10.1111/j.1600-0536.1995.tb00554.x)
- Romaguera C, Grimalt F, Vilaplana J. Nickel dermatitis from an infusion needle. *Contact Dermatitis*. 1985;12(3):181-182. doi:[10.1111/j.1600-0536.1985.tb01098.x](https://doi.org/10.1111/j.1600-0536.1985.tb01098.x)

7. van den Hove J, Jacobs MC, Tennstedt D, Lachapelle JM. Allergic contact dermatitis from acrylates in insulin pump infusion sets. *Contact Dermatitis*. 1996;35(2):108. doi:10.1111/j.1600-0536.1996.tb02305.x
8. Corazza M, Maranini C, Aleotti A, Virgili A. Nickel contact dermatitis due to the needle of an infusion pump, confirmed by microanalysis. *Contact Dermatitis*. 1998;39(3):144. doi:10.1111/j.1600-0536.1998.tb05875.x
9. Saccabusi S, Boatto G, Asproni B, Pau A. Sensitization to methyl methacrylate in the plastic catheter of an insulin pump infusion set. *Contact Dermatitis*. 2001;45(1):47-48. doi:10.1034/j.1600-0536.2001.045001047.x
10. Jolanki R, Kanerva L, Estlander T, Henriks-Eckerman ML, Suhonen R. Allergic contact dermatitis from phenoxyethoxy ethylacrylates in optical fiber coating, and glue in an insulin pump set. *Contact Dermatitis*. 2001;45(1):36-37. doi:10.1034/j.1600-0536.2001.045001036.x
11. Jolanki R, Estlander T, Suhonen R, Eckerman MLH, Kanerva L. Contact allergy to phenoxyethoxy ethyl acrylates. Poster presented during the 2nd congress of the European Society of Contact Dermatitis, Barcelona, Abstracts, p. 60. 1994.
12. Hamnerius N, Mowitz M. Intense skin reaction to a new glucose monitoring and insulin pump system. *Contact Dermatitis*. 2020;83(6):524-527. doi:10.1111/cod.13663
13. Ulriksdotter J, Svedman C, Bruze M, et al. Contact dermatitis caused by glucose sensors: 15 adult patients tested with a medical device patch test series. *Contact Dermatitis*. 2020;83(4):301-309. doi:10.1111/cod.13649
14. Mowitz M, Fornander L, Hosseiny S, Ryberg K, Bruze M. Patch testing with isobornyl acrylate in 16 Swedish patients with contact dermatitis from glucose sensors and/or insulin pumps. *Acta Derm Venereol*. 2019;99(13):1286-1287. doi:10.2340/00015555-3256
15. Ulriksdotter J, Svedman C, Bruze M, Mowitz M. Allergic contact dermatitis caused by dipropylene glycol diacrylate in the Omnipod® insulin pump. *Br J Dermatol*. 2022;186(2):334-340. doi:10.1111/bjd.20751
16. Ahrensboell-Friis U, Simonsen AB, Zachariae C, Thyssen JP, Johansen JD. Contact dermatitis caused by glucose sensors, insulin pumps, and tapes: results from a 5-year period. *Contact Dermatitis*. 2021;84(2):75-81. doi:10.1111/cod.13664
17. Cichoń M, Myśliwiec M, Trzeciak M. Role of acrylates in the development of contact dermatitis in diabetic patients: a Polish dermatology tertiary centre experience. *Contact Dermatitis*. 2024;90(2):126-133. doi:10.1111/cod.14436
18. Enberg J, Hamnerius N, Mowitz M. Allergic contact dermatitis caused by a new insulin pump system containing isobornyl acrylate. *Contact Dermatitis*. 2023;88(4):326-328. doi:10.1111/cod.14274
19. Cichoń M, Sokołowska-Wojdyło M, Trzeciak M. Allergic contact dermatitis elicited by insulin infusion sets: first case reported in Poland. *Contact Dermatitis*. 2023;88(5):404-406. doi:10.1111/cod.14281
20. Herman A, Darrigade AS, de Montjoye L, Baeck M. Contact dermatitis caused by glucose sensors in diabetic children. *Contact Dermatitis*. 2020;82(2):105-111. doi:10.1111/cod.13429
21. Herman A, Baeck M, de Montjoye L, et al. Allergic contact dermatitis caused by isobornyl acrylate in the EnLite glucose sensor and the paradigm MiniMed quick-set insulin infusion set. *Contact Dermatitis*. 2019;81(6):432-437. doi:10.1111/cod.13374
22. Raison-Peyron N, Mowitz M, Dereure O, Svedman C. Allergic contact dermatitis in a patient after five years of use of Omnipod insulin pump. *J Diabetes Sci Technol*. 2022;16(5):1342-1343. doi:10.1177/19322968221110578
23. Huang C, DeKoven J. An unexpected source of glucose monitor-associated allergic contact dermatitis. *Contact Dermatitis*. 2021;85:235-236. doi:10.1111/cod.13815
24. Oppel E, Högg C, Summer B, Ruëff F, Reichl FX, Kamann S. Isobornyl acrylate contained in the insulin patch pump OmniPod as the cause of severe allergic contact dermatitis. *Contact Dermatitis*. 2018;79(3):178-180. doi:10.1111/cod.13017
25. Kamann S, Aerts O, Heinemann L. Further evidence of severe allergic contact dermatitis from isobornyl acrylate while using a continuous glucose monitoring system. *J Diabetes Sci Technol*. 2018;12(3):630-633. doi:10.1177/1932296818762946
26. Klein A, Kamann S, Högg C, Oppel EM. Kontaktallergie auf kontinuierliche Glukosemesssysteme und insulin-patch-Pumpen [Contact allergy to continuous glucose monitoring systems and insulin patch pumps]. *Hautarzt*. German. 2020;71(Suppl 1):41-43. doi:10.1007/s00105-020-04623-1
27. Passanisi S, Lombardo F, Barbalace A, et al. Allergic contact dermatitis and diabetes medical devices: 2 clinical cases. *Contact Dermatitis*. 2018;79(2):115-117. doi:10.1111/cod.13012
28. Mowitz M, Hosseini S, Siemund I, Ulriksdotter J, Svedman C. Allergic contact dermatitis caused by the Dexcom G7 glucose sensor. *Contact Dermatitis*. 2024;90(5):495-500. doi:10.1111/cod.14514
29. Svedman C, Bruze M, Antelmi A, et al. Continuous glucose monitoring systems give contact dermatitis in children and adults despite efforts of providing less 'allergy-prone' devices: investigation and advice hampered by insufficient material for optimized patch test investigations. *J Eur Acad Dermatol Venereol*. 2021;35(3):730-737. doi:10.1111/jdv.16981
30. Svedman C, Ulriksdotter J, Lejdning T, Bruze M, Mowitz M. Changes in adhesive ingredients in continuous glucose monitoring systems may induce new contact allergy pattern. *Contact Dermatitis*. 2021;84(6):439-446. doi:10.1111/cod.13781
31. Mowitz M, Lejdning T, Ulriksdotter J, Antelmi A, Bruze M, Svedman C. Further evidence of allergic contact dermatitis caused by 2,2'-methylenebis(6-tert-butyl-4-methylphenol) monoacrylate, a new sensitizer in the Dexcom G6 glucose sensor. *Dermatitis*. 2022;33(4):287-292. doi:10.1097/DER.0000000000000767
32. Oppel E, Högg C, Oschmann A, Summer B, Kamann S. Contact allergy to the Dexcom G6 glucose monitoring system: role of 2,2'-methylenebis(6-tert-butyl-4-methylphenol) monoacrylate in the new adhesive. *Contact Dermatitis*. 2022;87(3):258-264. doi:10.1111/cod.14141
33. Hyry HSI, Liippo JP, Virtanen HM. Allergic contact dermatitis caused by glucose sensors in type 1 diabetes patients. *Contact Dermatitis*. 2019;81(3):161-166. doi:10.1111/cod.13337
34. Schwensen JF, Friis UF, Zachariae C, Johansen JD. Sensitization to cyanoacrylates caused by prolonged exposure to a glucose sensor set in a diabetic child. *Contact Dermatitis*. 2016;74(2):124-125. doi:10.1111/cod.12503
35. Gatica-Ortega ME, Mowitz M, Pastor-Nieto MA, et al. Contact dermatitis from glucose sensors in Spain: a multicentric approach. *Contact Dermatitis*. 2021;85(5):554-562. doi:10.1111/cod.13924
36. Oppel E, Kamann S, Reichl FX, Högg C. The Dexcom glucose monitoring system: an isobornyl acrylate-free alternative for diabetic patients. *Contact Dermatitis*. 2019;81(1):32-36. doi:10.1111/cod.13248
37. Mowitz M, Herman A, Baeck M, et al. N,N-dimethylacrylamide: a new sensitizer in the FreeStyle libre glucose sensor. *Contact Dermatitis*. 2019;81(1):27-31. doi:10.1111/cod.13243
38. Herman A, Mowitz M, Aerts O, et al. Unexpected positive patch test reactions to sesquiterpene lactones in patients sensitized to the glucose sensor FreeStyle Libre. *Contact Dermatitis*. 2019;81(5):354-367. doi:10.1111/cod.13330
39. Pyl J, Dendooven E, Van Eekelen I, den Brinker M, et al. Prevalence and prevention of contact dermatitis caused by FreeStyle Libre: a monocentric experience. *Diabetes Care*. 2020;43(4):918-920. doi:10.2337/dc19-1354
40. Kamann S, Heinemann L, Oppel E. Usage of hydrocolloid-based plasters in patients who have developed allergic contact dermatitis to isobornyl acrylate while using continuous glucose monitoring systems. *J Diabetes Sci Technol*. 2020;14(3):582-585. doi:10.1177/1932296819876964

41. Queirós CS, Alexandre MI, Garrido PM, Correia TE, Filipe PL. Allergic contact dermatitis to IBOA in FreeStyle Libre: experience from a tertiary care Portuguese hospital. *Contact Dermatitis*. 2020;83(2):154-157. doi:10.1111/cod.13578
42. Vidal-Albareda C, Yelmo-Valverde R, Solórzano-Zepeda C, Rodríguez-Muñoz N, de-la-Hoz- Caballer B, González-de-Olano D. Prevalence of contact dermatitis to glucose sensors in pediatric population and the main allergens involved. *Contact Dermatitis*. 2020;83(1):47-49. doi:10.1111/cod.13511
43. Herman A, de Montjoye L, Baeck M. Adverse cutaneous reaction to diabetic glucose sensors and insulin pumps: irritant contact dermatitis or allergic contact dermatitis? *Contact Dermatitis*. 2020;83:25-30. doi:10.1111/cod.1352930
44. Dendooven E, Foubert K, Goossens A, et al. Concomitant positive patch test reactions in FreeStyle-allergic patients sensitized to isobornyl acrylate. *Contact Dermatitis*. 2021;84(3):166-174. doi:10.1111/cod.13706
45. Dendooven E, Foubert K, Naessens T, Pieters L, Lambert J, Aerts O. Isocyanates may contribute to allergic contact dermatitis from diabetes devices and wound dressings. *Contact Dermatitis*. 2022;87(5):414-419. doi:10.1111/cod.14187
46. Alves da Silva C, Bregnhøj A, Mowitz M, Bruze M, Andersen KE, Sommerlund M. Contact dermatitis in children caused by diabetes devices. *Contact Dermatitis*. 2022;87(5):406-413. doi:10.1111/cod.14166
47. Oppel E, Kamann S, Heinemann L, Klein A, Reichl FX, Högg C. Free-style libre 2: the new isobornyl acrylate free generation. *Contact Dermatitis*. 2020;83(5):429-431. doi:10.1111/cod.13638
48. Seibold A. Minimizing adverse skin reactions to wearable continuous glucose monitoring sensors in patients with diabetes. *J Diabetes Sci Technol*. 2021;15(3):713-714. (Letter to the editor). doi:10.1177/1932296820984763
49. Dendooven E, Stappers S, Lambert J, Pieters L, Foubert K, Aerts O. Commonly problematic medical devices and associated contact allergens. *Allergy*. 2021;76(11):3517-3520. (Letter to the editor). doi:10.1111/all.14023
50. Peeters C, Herman A, Goossens A, Bruze M, Mowitz M, Baeck M. Allergic contact dermatitis caused by 2-ethyl cyanoacrylate contained in glucose sensor sets in two diabetic adults. *Contact Dermatitis*. 2017;77(6):426-429. doi:10.1111/cod.12873
51. Aschenbeck KA, Hylwa SA. A diabetic's allergy: ethyl cyanoacrylate in glucose sensor adhesive. *Dermatitis*. 2017;28(4):289-291. doi:10.1097/DER.0000000000000281
52. Lowe A, D'Souza N, Hughes J. A rarely reported contact allergy to ethyl cyanoacrylates in a continuous glucose sensor monitoring device in a child with diabetes, with a successful management outcome. *Br J Dermatol*. 2017;177(Suppl. 1):173.
53. Teufel-Schäfer U, Huhn C, Müller S, Müller C, Grünert SC. Severe allergic contact dermatitis to two different continuous glucose monitoring devices in a patient with glycogen storage disease type 9b. *Pediatr Dermatol*. 2021;38(5):1302-1304. doi:10.1111/pde.14767
54. Ng KL, Nixon RL, Grills C, Tam MM. Solution using Stomahesive® wafers for allergic contact dermatitis caused by isobornyl acrylate in glucose monitoring sensors. *Australas J Dermatol*. 2022;63(1):e56-e59. doi:10.1111/ajd.13675
55. Xiong M, Neeley AB. Adverse cutaneous reaction to the Eversense E3 continuous glucose monitor in a platinum-allergic patient. *Contact Dermatitis*. 2022;87(6):535-536. doi:10.1111/cod.14200
56. Joyanes Romo JB, Borja Segade JM, Moreno Lozano L, et al. Contact dermatitis caused by glucose sensors: more than allergy to isobornyl acrylate. *J Invest Allergol Clin Immunol*. 2022;32(3):232-233. doi:10.18176/jiaci.0739
57. Aerts O, Dendooven E, Raison-Peyron N. Sensitization to modified colophonium in glucose sensors: another problem for diabetes patients. *Contact Dermatitis*. 2022;87(6):553-555. doi:10.1111/cod.14213
58. Dendooven E, Dirinck E, Foubert K, Aerts O. "Re-testing" suggests that cosensitizations to isobornyl acrylate and sesquiterpene lactones may be due to cross-reactivity. *Contact Dermatitis*. 2022;86(1):57-59. doi:10.1111/cod.13972
59. Renaudin H, Darrigade AS, Dendooven E, Foubert K, Aerts O, Milpied B. Allergic contact dermatitis from a disposable blood pressure cuff containing isobornyl acrylate and 2-phenoxyethyl acrylate. *Contact Dermatitis*. 2021;84(6):462-464. doi:10.1111/cod.13761
60. Caldeira LE, Fernandes NP, Neto M, et al. A case of contact dermatitis caused by isobornyl acrylate in a glucose monitoring sensor. *Allergy: Eur J All Clin Immunol*. 2020;75(suppl 109):444-445. (Abstract).
61. DeKoven JG, Shear NH. Cutaneous contact allergy to a glucose monitor. *CMAJ*. 2020;192(11):E286. doi:10.1503/cmaj.191216
62. Tramontana M, Hansel K, Bianchi L, Agostinelli D, Stingeni L. Allergic contact dermatitis caused by a glucose monitoring system: an emerging side-effect of diabetes medical devices. *J Eur Acad Dermatol Venereol*. 2020;34(5):e223-e225. doi:10.1111/jdv.16178
63. Herman A, de Montjoye L, Marot L, Baeck M. Induction of leukoderma following allergic contact dermatitis to FreeStyle Libre. *Contact Dermatitis*. 2019;81(6):456-458. doi:10.1111/cod.13360
64. Jáuregui I, Sánchez J, Seguro A, Galán C, Seras Y, Gamboa PM. Allergic contact dermatitis by isobornyl acrylate in glucose monitoring devices: report of two cases. *Contact Dermatitis*. 2019;81(3):219-220. doi:10.1111/cod.13349
65. Shinkawa E, Washio K, Tatsuoka S, Fukunaga A, Sakaguchi K, Nishigori C. A case of contact dermatitis caused by isobornyl acrylate in FreeStyle Libre: the usefulness of film-forming agents. *Contact Dermatitis*. 2019;81(1):56-57. doi:10.1111/cod.13239
66. Wagner N, Elsbergen L, Feustel J. Isobornyl acrylate-contact allergen in continuous glucose sensor system. *Allergy: Eur J All Clin Immunol*. 2019;74(Suppl):502. (Abstract). doi:10.1111/all.13961
67. Mine Y, Urakami T, Matsuura D. Allergic contact dermatitis caused by isobornyl acrylate when using the FreeStyle® libre. *J Diabetes Investig*. 2019;10(5):1382-1384. doi:10.1111/jdi.13023
68. Kamann S, Oppel E. Hydrocolloid blister plaster decreases allergic contact dermatitis caused by Freestyle Libre and isobornyl acrylate. *Contact Dermatitis*. 2019;81(5):380-381. doi:10.1111/cod.13329
69. Fitzgerald M, Wakelin S, Bruze M. A sticky situation for diabetes: allergic contact dermatitis to isobornyl acrylate in continuous glucose-monitoring devices. *Br J Dermatol*. 2019;181(Suppl 1):158 (Abstract).
70. Corazza M, Scuderi V, Musmeci D, Foti C, Romita P, Borghi A. Allergic contact dermatitis caused by isobornyl acrylate in a young diabetic patient using a continuous glucose monitoring system (Freestyle Libre). *Contact Dermatitis*. 2018;79(5):320-321. doi:10.1111/cod.13075
71. Bulteel C, Huygens S, Gilissen L, Goossens A. Isobornyl acrylate in glucose sensors causes allergic contact dermatitis in children. *Pediatr Dermatol*. 2018;35(Suppl 2):S22 (Poster).
72. Lyngstadaas AV, Holm JØ, Krogvold L, Måløy AK, Ingvaldsen CA. A toddler with systemic contact dermatitis caused by diabetes devices. *Skin Health Dis*. 2023;3(4):e234. doi:10.1002/ski2.234
73. Persad S, Lejdning T, Svedman C, Mowitz M. The missing link: allergic contact dermatitis caused by 1,6-hexanediol diacrylate in a glucose sensor transmitter. *Contact Dermatitis*. 2023;89(4):303-305. doi:10.1111/cod.14385
74. Gatica-Ortega ME, Pastor-Nieto MA, Torres-Aranda R, Luna-Bastante L, Pérez-Hortet C. Contact dermatitis from MiaoMiao (an add-on transmitter to FreeStyle Libre flash glucose monitoring system). *Contact Dermatitis*. 2021;85(1):92-94. doi:10.1111/cod.13788
75. Aerts O, Herman A, Mowitz M, Bruze M, Goossens A. Isobornyl acrylate. *Dermatitis*. 2020;31(1):4-12. doi:10.1097/DER.0000000000000549

76. Bembry R, Brys AK, Atwater AR. Medical device contact allergy: glucose monitors and insulin pumps. *Curr Derm Rep*. 2022;11:13-20. doi:10.1007/s13671-021-00352-3
77. Navarro-Triviño FJ. Skin reactions to glucose sensors: present and future. *Actas Dermosifiliogr (Engl Ed)*. 2021;112(5):389-391. English, Spanish. doi:10.1016/j.ad.2020.09.008
78. Steunebrink IM, de Groot A, Rustemeyer T. Contact allergy to acrylate-containing nail cosmetics: a retrospective 8-year study. *Contact Dermatitis*. 2024;90:262-265. doi:10.1111/cod.14475
79. Aerts O, Herman A, Bruze M, Goossens A, Mowitz M. FreeStyle Libre: contact irritation versus contact allergy. *Lancet*. 2017;390(10103):1644. doi:10.1016/S0140-6736(17)32142-6
80. Ulriksdotter J, Mowitz M, Svedman C, Bruze M. Patch testing and diagnosis when suspecting allergic contact dermatitis from medical devices. *Contact Dermatitis*. 2020;83(4):333-335. doi:10.1111/cod.13650
81. Hamada H, Bruze M, Zimerson E, Isaksson M, Engfeldt M. Sensitization and cross-reactivity patterns of contact allergy to diisocyanates and corresponding amines: investigation of diphenylmethane-4,4'-diisocyanate, diphenylmethane-4,4'-diamine, dicyclohexylmethane-4,4'-diisocyanate, and dicyclohexylmethane-4,4'-diamine. *Contact Dermatitis*. 2017;77(4):231-241. doi:10.1111/cod.12809
82. Andersen KE, White IR, Goossens A. Chapter 31: Allergens from the European baseline series. In: Johansen JD, Frosch PJ, Lepoittevin J-P, eds. *Contact Dermatitis*. 5th ed. Springer-Verlag; 2011:545-590.
83. Downs AM, Sansom JE. Colophony allergy: a review. *Contact Dermatitis*. 1999;41(6):305-310. doi:10.1111/j.1600-0536.1999.tb06178.x
84. Steunebrink IM, de Groot A, Rustemeyer T. Presence of 2-hydroxyethyl methacrylate (HEMA) and other (meth)acrylates in nail cosmetics, and compliance with EU legislation: an online market survey. *Contact Dermatitis*. 2024;90(1):60-65. doi:10.1111/cod.14441
85. Bruze M, Trulsson L, Bendsøe N. Patch testing with ultrasonic bath extracts. *Am J Cont Derm*. 1992;3(3):133-137. doi:10.1053/ajcd.1992.2413
86. Bruze M. The use of ultrasonic bath extracts in the diagnostics of contact allergy and allergic contact dermatitis. In: Lachapelle J-M, Bruze M, Elsner P, eds. *Patch Testing Tips*. Springer; 2014:129-142.
87. Bolinder J, Antuna R, Geelhoed-Duijvestijn P, Kröger J, Weitgasser R. Novel glucose-sensing technology and hypoglycaemia in type 1 diabetes: a multicentre, non-masked, randomised controlled trial. *Lancet*. 2016;388(10057):2254-2263. doi:10.1016/S0140-6736(16)31535-5
88. Bolinder J, Antuna R, Geelhoed-Duijvestijn P, Kröger J, Weitgasser R. Cutaneous adverse events related to FreeStyle Libre device – Authors' reply. *Lancet*. 2017;389(10077):1396-1397. doi:10.1016/S0140-6736(17)30893-0
89. Lombardo F, Salzano G, Crisafulli G, et al. Allergic contact dermatitis in pediatric patients with type 1 diabetes: an emerging issue. *Diabetes Res Clin Pract*. 2020;162:108089. doi:10.1016/j.diabres.2020.108089
90. Lombardo F, Passanisi S, Caminiti L, et al. High prevalence of skin reactions among pediatric patients with type 1 diabetes using new technologies: the alarming role of colophonium. *Diabetes Technol Ther*. 2020;22(1):53-56. doi:10.1089/dia.2019.0236
91. Medicines and Healthcare products Regulatory Agency. FreeStyle Libre flash glucose sensor – Use of barrier methods to reduce skin reactions to the sensor adhesive (MDA/2019/003). Accessed March 17, 2024 <https://www.gov.uk/drug-device-alerts/freestyle-libre-flash-glucose-sensor-use-of-barrier-methods-to-reduce-skin-reactions-to-the-sensor-adhesivemda-2019-003>
92. Paret M, Barash G, Rachmiel M. "Out of the box" solution for skin problems due to glucose-monitoring technology in youth with type 1 diabetes: real-life experience with fluticasone spray. *Acta Diabetol*. 2020;57(4):419-424. doi:10.1007/s00592-019-01446-y
93. Cressey BD, Belum VR, Scheinman P, et al. Stoma care products represent a common and previously underreported source of peristomal contact dermatitis. *Contact Dermatitis*. 2017;76:27-33. doi:10.1111/cod.12678
94. Navarro-Triviño FJ, Sierra Cerdán MJ, Ruiz-Villaverde R, Tegaderm TMIV. Advanced can prevent eczematous reaction caused by Dexcom glucose monitoring system. *Int J Dermatol*. 2020;59(5):e166-e168. doi:10.1111/ijd.14772
95. Pereira TM, Flour M, Goossens A. Allergic contact dermatitis from modified colophonium in wound dressings. *Contact Dermatitis*. 2007;56(1):5-9. doi:10.1111/j.1600-0536.2007.01026.x
96. Körber A, Kohaus S, Geisheimer M, Grabbe S, Dissemmond J. Allergisches Kontaktekzem durch einen Hydrokolloidverband bei Sensibilisierung auf Kolofonium [Allergic contact dermatitis from a hydrocolloid dressing due to colophony sensitization]. *Hautarzt*. 2006;57(3):242-245. German. doi:10.1007/s00105-005-0913-x
97. Herman A, Uter W, Rustemeyer T, et al. Position statement: the need for EU legislation to require disclosure and labelling of the composition of medical devices. *J Eur Acad Dermatol Venereol*. 2021;35(7):1444-1448. doi:10.1111/jdv.17238
98. Herman A, de Montjoye L, Tromme I, Goossens A, Baeck M. Allergic contact dermatitis caused by medical devices for diabetes patients: a review. *Contact Dermatitis*. 2018;79(6):331-335. doi:10.1111/cod.13120
99. Herman A, Goossens A. The need to disclose the composition of medical devices at the European level. *Contact Dermatitis*. 2019;81(3):159-160. doi:10.1111/cod.13354

SUPPORTING INFORMATION

Additional supporting information can be found online in the Supporting Information section at the end of this article.

How to cite this article: de Groot A, van Oers EM, Ipenburg NA, Rustemeyer T. Allergic contact dermatitis caused by glucose sensors and insulin pumps: A full review. *Contact Dermatitis*. 2024;1-12. doi:10.1111/cod.14697